



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°R60/1991-NL-97.15
Project number 10067238
Page 1 of 2

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI Certin B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: A.J. Nederlof

Applicant

Name: A & D Instruments Ltd
Address: Abingdon Science Park,
Adingdon, Oxford,
OX14 3YS United Kingdom

Manufacturer of the certified pattern

Name: A & D Instruments Ltd
Address: Abingdon Science Park,
Adingdon, Oxford,
OX14 3YS United Kingdom

Identification of the certified pattern

Type : LC-5207
 $P_1 = 0.7$

Maximum Capacity (E_{max})	20 t and 25 t
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Ratio of the minimum LC verification interval ($Y = E_{max} / V_{min}$)	4000* or 6000**

- * Temperature range: -5 °C ./ 40 °C and maximum excitation 15 V AC/DC.
** Temperature range: -10 °C ./ 40 °C and maximum excitation 15 VDC.

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht (NL)
Telephone +31 78 6332332
Telefax +31 78 6332309

NMI B.V. (Registered at the Chamber of Commerce
Delft number 26028701)

Subsidiary companies:
NMI Certin B.V. (26033418)
NMI Van Swinden Laboratorium B.V. (26028703)
NMI Inspecties en Kansspeltechniek B.V. (26028700)
NMI International B.V. (26039176)

This certificate is issued under the provision that
NMI B.V. nor its subsidiary companies accept any
liability.

Reproduction of the complete certificate is allowed.
Parts of the certificate may only be reproduced
after written permission.



**QUALIFIED
BY STERLAB**
Reg. nr. L 029



Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°R60/1991-NL-97.15
Project number 10067238
Page 2 of 2

This certificate attests the conformity of the above-mentioned pattern (represented by the samples identified in the associated test report, the test certificate and the description with number TC2965 and the appertaining documentation folder, with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

R60
edition 1991 (E)
for accuracy class C

This certificate relates only to the metrological and technical characteristics of the pattern of the Loadcell concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.

The conformity was established by tests described in the associated test report N° R60/1991-NL-97.15, that includes 37 pages.

The issuing authority
A.J. Nederlof

14 April 1997

The OIML member
G.J. Faber

14 April 1997

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.