

Nederlands Meetinstituut

Member State
The Netherlands

OIML Certificate N°
R76/1992-NL-93.16
Revision 1

OIML CERTIFICATE OF CONFORMITY

Issuing authority

Name: NMI IJkezen B.V.
Address: Hugo de Grootplein 1, Dordrecht
Person responsible: Mr. M. Charité

Applicant

Name: A&D Instruments Ltd
Address: Abingdon Science Park
Abingdon, Oxford
OX14 3YS United Kingdom

Manufacturer of the certified pattern


Name: A&D
Address: Abingdon Science Park
Abingdon, Oxford
OX14 3YS United Kingdom

Identification of the certified pattern

Type FS
Maximum weighing capacity $6 \text{ kg} \leq \text{Max} \leq 30 \text{ kg}$ or the equivalent
in metric or imperial pounds.
Maximum of two partial weighing ranges for multi interval scales.
 $T \leq - \text{Max}$
 $n \leq 3000$ divisions
Operating temperature $-10 \text{ °C} / +40 \text{ °C}$

This certificate attests the conformity of the above-mentioned pattern (represented by the sample identified in the associated test report, the EC type-approval certificate number T2290 and the description number T2290-1 and the documentation folder number T2290-2, both appertaining to this EC type-approval certificate) with the requirements of the following Recommendation(s) of the International Organization of Legal Metrology (OIML):

R76
edition 1992
for accuracy class **III**



Nederlands Meetinstituut

OIML Certificate N°
R76/1992-NL-93.16
Revision 1

This certificate relates only to the metrological and technical characteristics of the pattern of the instrument concerned, as covered by the relevant OIML International Recommendation(s).

This certificate does not bestow any form of legal international approval.
The conformity was established by tests described in the associated test report
N° R76/1992-NL-93.16, that includes 59 pages.

The issuing authority
M. Charité

21 December 1993

The OIML member
G.J. Faber

21 December 1993

*
* *

Important note: Apart from the mention of the certificate's reference number and the name of the OIML Member State in which the certificate was issued, partial quotation of the certificate or of the associated test report is not permitted, though they may be reproduced in full.